



VISGENEER INC.

Address: No.188,Sec. 3,Gongdao 5th Rd.,Hsinchu City 30069, Taiwan

Tel: +886-3-5160111 Fax: +886-3-5160161 Web: www.visgeneer.com.tw

TO WHOM IT MAY CONCERN:

We have performed eBsensor II (C740) glucose meter test with comparison to Roche (C946), One Touch (C25) and YSI in 7 different blood glucose levels. The results are stated below:

YSI	eBsensor II (C740)	Roche (C946)	One Touch(C25)
55	58	50	50
112	110	115	115
160	159	166	161
262	259	251	264
370	369	349	368
461	469	467	457
570	577	561	537

Test By: Tina

Review By: Angel



VISGENEER INC.
NO. 188, Sec. 3, Gongdao 5th Rd.,
Hsinchu City 30069, Taiwan
Tel: 886-3-5160111
Fax: 886-3-5160161
<http://www.visgeneer.com.tw>



VISGENEER INC.
NO. 188, Sec. 3, Gongdao 5th Rd.,
Hsinchu City 30069, Taiwan
Tel: 886-3-5160111
Fax: 886-3-5160161
<http://www.visgeneer.com.tw>